EQUIPMENT REQUIREMENTS

As per Revised BDS Course Regulations, 2007

Department: Oral Medicine and Radiology

		Qu	antity	Available	Adequate	Remark
Name	Specification	50 Intake	100 Intake	Quantity	/Inadequate te	Kemark
	(Space allocation for each dental chair 100 Sq.Ft. in All the Departments)					
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventry Suction, X-Ray viewer, 3 way syringe, instrument tray, Dental Operator's Stool with height adjustment		12	12	Adequate	
Panoramic X-Ray with Cephalometric	Preferably digital		1	1	Adequate	
Intra Oral X-ray Unit	70 KV, 8 mA, high frequency preferably Digital timer		3	3	Adequate	
Pulp testers	Digital				Adequate	
Automatic periapicalX- ray Developer			1	1	Adequate	
Automatic Panoramic with Cephalometric X- ray Developer			1	No		
X-ray viewer	For Panoramic and Cephalometric films		2	2	Adequate	
Radiovisiography	Digital Intra X-ray system with one Sensor andSoftware		1	1	Adequate	
GeneralX-ray Unit			1	1	Adequate	100 miles (100 miles (
Ortho Pantmograph	Preferably digital		1	1	Adequate	
AutomaticDevelopers / Dark Room			2	2	Adequate	
Lead aprons			2	2	Adequate	
Lead Gloves			1	1	Adequate	
X-ray Hangers			6	6	Adequate	
X-ray Viewers			2	2	Adequate	
Diagnostic Kits	Mouth mirror, dental probe, college tweezers		40	40	Adequate	
Lead Screen			1	1	Adequate	
Biopsy Kit			1	1	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		2	2	Adequate	
Computers	Minimum Pentium – IV		1	1	Adequate	2700 90
Intra Oral Camera	With High resolution		1	1	Adequate	
Needle Burner with Syringe Cutter			2	2	Adequate	



Department: Oral Pathology and Oral Microbiology

Name		Quantity		Available	Adaguata	Remark
	Specification	Intake Intake	Quantity	Adequate /Inadequat e	Kemark	
Dental Chairs and Units	With shadowless lamp, spittoon, 3 way syringe, instrument tray and suction		2	2	Adequate	
Microscopes			40	40	Adequate	17.128
Microtome			1	1	Adequate	
Wax bath			1	1	Adequate	
Water bath		88	1	1	Adequate	
Knife sharpener			1	1	Adequate	1
Hot plate	2		1	1	Adequate	
Spencer Knife			1	1	Adequate	

Departme	nt: Public Health Dentistry				1	1 00 - 100
And the			ntity	Available	Adequate	Remark
Name	Specification	50 Intake	100 Intake	Quantity	/Inadequat	
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, micromotor control light cure 3 way syringe, X-ray viewer, instrument tray Dental Operators stool with height adjustment with shadowless lamp, spittoon, 3 way syringe, instrument tray and suction, micromotor, airotor, light cure		16	16	Adequate	
Autoclaves	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		2	2	Adequate	
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket with digital timer	47.54	2	2	Adequate	107 107
Needle burner with syringe cutter			4	4	Adequate	
MOBILE CLINIC					Adequate	
Mobile dental van	Mobile dental van with two dental chairs with all the attachments and adequate sitting space for 15 to 20 people				Adequate	
Dental chair with unit	Hydraulically operated with spittoon attachment, halogen light with 2 intensity, air ventury suction, airrotor, micromotor, 3 way-scaler and light cure, x-ray viewer, instrument tray, operating stool				Adequate	
Autoclaves	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		1	1	Adequate	
Intraoral x-ray	Portable, 70 KV, 8 mA		1	1	Adequate	
Glass bead sterilisers			1	1	Adequate	
Compressor	1.25 HP		1.	1	Adequate	
Metal Cabinet	With wash basin		1	1	Adequate	
Portabledental chair	Suitcase unit with airotor, micromotor, scaler and compressor 0.25 HP		2	2	Adequate	
Stabiliser	4 KV		1	1	Adequate	
Generator	4 KV		1	1	Adequate	
Water tank	400 ltrs		1	1	1	
Oxygen cylinder	SOF * 'MAR' *		1	Yes	1	

AHOPAN JAYAONS

Publicaddress system	1	Yes	Adequate	
TV and video cassette player	1	1	1	Adequate
Demonstration models	1	1	1	Adequate

2 2 2 2 3 3 3 3 3 3 3 3 3 3		Qua		Available	Adequat	Remark
Name	Specification	50	100	Quantity	е	
		Intake	Intake		/Inadequ	
					ate	
Dental Chairs and	Electrically operated, Spittoon attachment, Halogen		20	20	Adequate	
Jnits	Light with 2 intensity, high power motorized					
,,,,,	evacuation system, Air rotor with miniature, Airrotor					
	HPS, Micro motor with miniature contrangle Hand					
	piece, 3 way syringe, ultrasonic sealer with 3 tips,					
	Light cure unit LED based heat free, X-ray viewer,					
	instrument tray Dental Operator's stool with height					
	adjustment (Pedo chair preferred)					
Autoclaves	Having wet and dry cycle, which can achieve 135°C		2	2	Adequate	2014
	with minimum capacity of 20 liters		_		VT)	
Iltrasonic cleaner	Minimum capacity 13 liters with mesh bucket with	1	2	2	Adequate	
recusormo organier	digital timer		_	_		
Needle Burner			3	3	Adequate	
With syringe					Proceedings of the Control of the Co	
cutter						
Amalgamator			1	1	Adequate	
Pulp Tester Digital			1	1	Adequate	
Rubber dam kit	1		5	5	Adequate	
or pedo			,	3	lacquete	
Apex locator			1	1	Adequate	
indo motor	With torque control HPs		1	1	Adequate	
njectable gutta	With torque control in 3		1	1	Adequate	
percha with			1	1 2	lacquate	
condensation						
Radiovisiography	Digital intra X-ray system with pedo sensor and		1	1	Adequate	
iddio visio Brapily	software		-	+	Jucquate	
ntra Oral Camera	With high resolution		1	1	Adequate	
Scale instruments	Withinghiesoldton		10	10	Adequate	
Restorative		*	10	Language Committee Committ	Adequate	
nstruments			10	10	Adequate	
extraction forceps		n - n a -)	10	10	Adequate	
ntra-oral X-ray		-	1	10	Adequate	
Automatic			1	1	Adequate	
Developer			+	+	Aucquate	
Computer	Minimum Pentium – IV		1	1	Adequate	10 - 10 - 10 - 10 - 10 - 10 - 10 - 10 -
computer	PEDO LAB		<u> </u>		Adequate	W
Plaster dispenser	One each for plaster and stone plaster		2	2	Adequate	
Model Trimmer	With diamond disc		1	1	Adequate	
Model Trimmer	Double disc one diamond and one carborandam disc		1	1	Adequate	
Welder with			1		Adequate	
			1	1	Auequate	
soldering attachments			48	4		
	OTEGE * ADOP				Adaguata	
/ibrator	(3)		2	2 23/0	Adequate	
	OUTEGE * AUATOR SOLVE OF SOLVE					

Lab micro motor	Heavy duty with Hps	3	3	Adequate
Dental Lathe		1	1	Adequate
Model Trimmer		1	1	Adequate
¥	d Production			
Steam cleaner		1	1	Adequate ·
Pressure moulding Machine		1	1	Adequate
Carborandum Disc	BU = SOPAL	1	1	Adequate
Diamond disc		1	1	Adequate

		Qua	ntity	Available	Adagust	Remark
Name	Specification	50 Intake	100 Intake	Quantity	Adequat e /Inadequ ate	Kemark
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, micromotor control light cure 3 way syringe, X-ray viewer, instrument tray Dental Operators stool with height adjustment		18	18	Adequate	
Unit mount scaler			5	5	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		2	2	Adequate	
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket with digital timer		2	2	Adequate	
ORTHO LAB		- XXX-XXX			Adequate	
Plaster dispenser	One each for plaster and stone plaster		2	2	Adequate	
Vibrator	191. A. 3		2	2	Adequate	
Model Trimmer			2	2	Adequate	
Micromotor	Heavy duey		4	4	Adequate	
Lathe			2	2	Adequate	
X-ray viewers		1000	2	2	Adequate	
OPG with cephalostat	If available in radiology its is adequate		1	1	Adequate	
Welders	2		4	4	Adequate	
Blue Torch			1	1	Adequate	
Base Formers			4	4	Adequate	
Typodont			4	4	Adequate	
Set of pliers			10	10	Adequate	
Welder with soldering attachments		4	1	1	Adequate	
Hydro solder			1	1	Adequate	
Typhodont articulator	With metal teeth wax rim of Class I, II, III		4	4	Adequate	
Pressure moulding machine			1	1		



Department: Periodontology

		Qua	ntity	Available	Adequat	Remark
Name	Specification	50 Intake	100 Intake	Quantity	e /Inadequ ate	
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, with shadowless lamp, spittoon, 3 way syringe, Ultrasonic sealer with 3 tips, X-ray viewer, instrument tray Dental Operators stool with height adjustment		34	34	Adequate	
Scaling Instrument sets			10	10	Adequate	
Surgical instrument sets			6	6	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		3	3	Adequate	20000000
Ultrasonic scaler	Minimum capacity 13 liters with mesh bucket		2	2	Adequate	
Electrosurgical Cautery			1	1	Adequate	
Needle burner With syringe cutter			6	6	Adequate	
LASER	Soft tissue laser		1	1	Adequate	
Surgical motor With physio dispenser			1	1	Adequate	

Department: Conservative Dentistry and Endodontics

		Qua	ntity	Available	Adequat	Remark
Name	Specification	50 Intake	100 Intake	Quantity	e /Inadequ ate	Kemark
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, with shadowless lamp, spittoon, 3 way syringe, Ultrasonic sealer with 3 tips, X-ray viewer, instrument tray Dental Operators stool with height adjustment		34	34	Adequate	
Rubber dam kits			6	6	Adequate	
Restorative instruments kits			10	10	Adequate	
R.C.T. instrument kits			10	10	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		3	3	Adequate	
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket		2	2	Adequate	
NeedleBurner With Syringe Cutter			4	4	Adequate	
Amalgamator	With auto proportion, auto dispenser	21	3	3	Adequate	
Rubber dam kits	GENOATAL OPAO		6	6	Service Man	e entresso medic

Pulp Tester Digital		4	4	Adequate	
Apex locator		2	2	Adequate	SUIS
Glass bead		6	6	Adequate	
sterilizers				Adequate	
Plaster dispenser		2	2	Adequate	
Vibrator		2	2	Adequate	
Ceramic Unit		1	1	Adequate	
Casting machine		1	1	Adequate	
	Proper radiation safety	1	1	Adequate	
Automatic Developer		1	1	Adequate	
Radiovisiography	RVG with Computer	1	1	Adequate	
Endo motor	With torque control Hps	1	1	Adequate	37. 15.
Bleaching Unit		1	1	Adequate	
Magnification loops		2	2	Adequate	
Injectabl gutta percha		2	2	Adequate	
PHANTOM LAB UNIT	Phantom Table fitted with Halogen Operating Light Phantom Head body type neck joint for all the movement, TMJ movement, Modular with Air rotor, Micro motor with contra angle Hps, 3 way syringe, jaw with ivorine teeth, preferably soft gingival, dental operator's stool (not to use extracted or cadaver teeth).	60	60	Adequate	

Chemical Laboratory

		Qua	ntity	Available	Adequat	Remark
Name	Specification	50 Intake	100 Intake	Quantity	e /Inadequ ate	Kemark
Plaster Dispenser	One each for plaster and stone plaster		2	2	Adequate	
Model Trimmer	Carborandum disc		1	1	Adequate	
	Diamond disc		1	1	Adequate	
Lathe	Heavy Duty		2	2	Adequate	
Lab Micromotor	With heavy duty handpiece		3	· 3	Adequate	
Ultrasonic cleaner	Minimum capacity 5 litres		1	1	Adequate	
Spindle Grinder			1	1	Adequate	
Vibrator			2	2	Adequate	
Burnout furnace			1	1	Adequate	
Porcelain furnace			1	1	Adequate	
Sandblasting Machine			1	1	Adequate	
Lab Airotor			1	1	Adequate	
Pindex System		n.	1	1	Adequate	
Circular saw			1	1	Adequate	
Vacuum mixer			1	1	Adequate	
Pnematic chisel			1	1	Adequate	
Casting machine	Motor cast / induction casting preferred		1	1		

Department: Oral & Maxillofacial Surgery

A) Exodontia

		Qua	ntity	Available	Adequate	Remark
Name	Specification	50 Intake	100 Intake	Quantity	/Inadequate e	
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, X-ray viewer, 3 way syringe, instrument tray Dental Operators stool and height adjustment and suction, Micromotor / Airotor		30	30	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		2	2	Adequate	
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket		2	2	Adequate	
Needle Burner with Syringe Cutter			6	6	Adequate	===44 F
Extraction forceps sets	Complete set		20	20	Adequate	
Dental elevators	Complete set		10	10	Adequate	
Minor Oral Surgery kits			6	6	Adequate	
Emergency drugs tray			1	1	Adequate	
Oxygen cylinder with mask			1	1	Adequate	
X-ray viewers			2	2	Adequate	
Computer	Minimum Pentium – IV		1	Yes		

B) Minor Surgery

			Quantity		Adamı	Remark
Name	Specification	50 Intake	100 Intake	Quantity	Adequ ate /Inadeq uate	Kemark
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evaculation system, Air ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contra-angle Handpiece, 3 way syringe, instrument tray Dental Operators stool and height adjustment and suction.		5	5	Adequate	

Department: Prosthodontics and Crown & Bridge

		Quantity		Available	A -1	Remark
Name	Specification	50 Intake	100 Intake	Quantity	Adequat	Kemark
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, 3 way syringe, Ultrasonic sealer with 3 tips, Light cure unit, instrument tray and suction, Dental Operators stool with height adjustment	*	34	34	Adequate	
Semi-adjustable articulator	With face bow		2	2	Adequate	

- V	<u> </u>		-	To de sucreta
Extra oral / intra		2	2	Adequate
oral tracer		2	2	Adequate
Dewaxing unit				Adequate
Curing unit	N Land of the land	2	2	Adequate
Dentalcasting		1	1	Adequate
machine			4	Adequate
Wax burnout		1	1	Adequate
Furnace			1	Adequate
Preheating		1	1	Adequate
furnace		2	2	Adequate
Surveying unit		4	4	Adequate
Heavy duty hand Piece		4		
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	2	2	Adequate
NeedleBurner		2	2	Adequate
With Syringe				
Cutter			desc.	
Plaster dispenser	One each for plaster and stone plaster	2	2	Adequate
ModelTrimmer		1	1	Adequate
with				
carborandum				
Disc	1 7		14.1	
Model Trimmer		2	2	Adequate
with diamond				
Disc				
Acryliser		3	3	Adequate
Lathe		2	2	Adequate
Flask press	2014-01 89/000 89/01 898-00 11/04 12/01	4	4	Adequate
Deflasking unit		4	4	Adequate
Dewaxing unit		3	3	Adequate
Hydraulic Press		3	3	Adequate
Mechanical Press		2	2	Adequate
Vacuum mixing	(F-24): 3	1	1	Adequate
machine				
Lab Micro motor	With heavy duty handpiece	4	4	Adequate
Curing pressure		1	1	Adequate
Pot				
Porcelain furnace		2	2	Adequate
Vibrator		2	2	Adequate
Sandblasting unit		2	2	Adequate
Ultrasonic cleaner		2	2	Adequate
Model Trimmer		4	4	Adequate
Hot water		2	2	Adequate
sterilizer				10
Geyser	Compound bath	2	2	Adequate
H.P. grinder with		3	3	Adequate
suction				
Heavy duty lathe		2	2	Adequate
Phantom heads		50	50	Adequate
Pre-clinical	Gas connection & Bunsen burner	100	100	Adequate
working tables				



Ceramic and Cast Partial Laboratory

laster and stone	50 Intake	100 Intake 2 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	Quantity 2 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	e /Inadequ ate Adequate	
		1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	Adequate	
		1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	Adequate	
ainers		1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	Adequate	
ainers		1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	Adequate	
ainers		1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1 1 1 1	Adequate	
ainers		1 1 1 1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1 1 1	Adequate	
ainers		1 1 1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1 1	Adequate	
		1 1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1 1	Adequate	
		1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1 1	Adequate	
		1 1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1 1	Adequate	
		1 1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate	
		1 1 1 1 1 1 1 1	1 1 1 1 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate	
		1 1 1 1 1 1 1 1	1 1 1 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate	
		1 1 1 1 1 1 1 1	1 1 1 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate	
		1 1 1 1 1 1 1 1	1 1 1 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate	
		1 1 1 1 1 1	1 1 1 1 1 1	Adequate Adequate Adequate Adequate Adequate	
		1 1 1 1 1 1	1 1 1 1 1 1	Adequate Adequate Adequate Adequate Adequate	
		1 1 1 1 1	1 1 1 1	Adequate Adequate Adequate Adequate	
		1 1 1 1	1 1 1 1	Adequate Adequate Adequate	
		1 1 1	1 1 1	Adequate Adequate	
		1 1 1	1 1 1	Adequate Adequate	
		1 1	1 1	Adequate	27.
		1	1	The second secon	
	-	100			
		1 L	1 1	Adequate	
The second secon		1	1 1	Adequate	
		1	1 1	Adequate	
У		1	1 1	Adequate	-
		1	1	Adequate	
		288			-
o proforabl		0.00		Textures on the street come	
nless steel fitted nsen burner, ai ng stool. nber of lab micro	d r	20	20	Adequate	
				Adequate	
laster and stone	1	2	2	Adequate	
	Î	2	2	Adequate	
and the second second	11.00	2	2	Adequate	
	A44	1	1	Adequate	
		1	1	Adequate	
3.7 - 1500		1	1	Adequate	
		1			l/
-univ.			155.00		
	nless steel fitted in the sen burner, air gradual stool. It is attached in the sen burner attached in the sen burner, and the sen burner attached in the sen burner, and the sen burner attached in the sen burner a	aless steel fitted asen burner, air ag stool. ber of lab micro attached hand aster and stone	aster and stone 20 21 22 21 11 11	e preferably 60 60 less steel fitted nsen burner, air g stool. ber of lab micro attached hand 22 2 2 2 2 1 1 1 1 1 1 1 1 1 1 1 1 1 1	e preferably filess steel fitted first burner, air files steel fitted filess steel fitted files burner, air files stool. Ser of lab micro attached hand Adequate aster and stone 2 2 Adequate 2 Adequate 2 2 Adequate 1 1 Adequate

EQUIPMENT REQUIREMENTS

Departmen: Prosthodontics and Crown & Bridge

Name	Specification	Quantity	Availab le Quantit	Adequate	Remark
Electrical Dental Chairs and Units	spittoon, 3 way syringe instrument tray and motorized suction,	, One Chair and t, unit per Po d student and two chairs with uni r for the faculty			
		1 Unit 2 Units	Yes	Adequate	
Articulators – semi adjustable/ adjustable with face bow		6	6	Adequate	
Micromotor – (Lab Type can also be attached (fixed) to wall		2	2	Adequate	
Ultrasonic scaler		2	2	Adequate	
Light cures		2	2	Adequate	
Hot air oven		1	1	Adequate	
Autoclave		2	2	Adequate	
Surveyor		2	2	Adequate	
Refrigerator	****	1	1	Adequate	102
K-ray viewer		1	1	Adequate	
Pneumatic, Crown bridge remover		2	2	Adequate	
Needle destroyer		1	1	Adequate	
ntra oral camera		1	1	Adequate	
Digital SLR camera		1	1	Adequate	
Computer with internet connection with		1	1	Adequate	
attached printer and scanner					
_CD projector	***************************************	1	1		
Clinical Lab for Prosthe	tics				l
Plaster dispenser		2	2	Adequate	
Model trimmer with carborandum Disc		1	1	Adequate	
Model trimmer with diamond disc		1	1	Adequate	
High speed lathe		2	2	Adequate	G)
/ibrator		2		Adequate	
Acrylizer		1		Adequate	
Dewaxing unit	,	1		Adequate	
Hydraulic press		1		Adequate	
Mechanical press		1		Adequate	
Vacuum mixing machine		1		Adequate	
Micro motor lab type		2		Adequate	
Curing pressure pot		1 1		Adequate	
Pressure molding machine		1 1		Adequate	
Chrome – Cobalt Lab Equip	ment			- Was discussed	
Duplicator		1	1	Adequate	
Pindex system		$+$ $\frac{1}{1}$ $+$ $-$		Adequate	16-27
Burn-out furnace		1		Adequate	
Welder		1	1	Adequate	atanawy via tanan
Sandblaster	Micro and macro	1		Adequate	
Electro – polisher		1		Adequate	
Model trimmer with carborandum disc	OFENDAYAL UP TO	1		Adequate	

1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	Adequate
1 1 1 1 1 2 2 2 1 1 1 1	Adequate
1 1 1 1 1 2 2 2 1 1 1 1	Adequate
1 1 1 1 1 2 2 2 1 1 1 1	Adequate
1 1 1 1 1 2 2 2 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate
1 1 1 1 1 2 2 2 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate
1 1 1 1 1 2 2 2 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate
1 1 1 1 2 2 2 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate Adequate
1 1 1 2 2 2 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate
1 1 1 2 2 2 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate
1 1 1 2 2 2 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate Adequate
1 1 2 2 2 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate
1 1 2 2 2 1 1 1 1	Adequate Adequate Adequate Adequate Adequate Adequate
1 2 2 1 1 1 1	Adequate Adequate Adequate Adequate
2 2 1 1 1 1	Adequate Adequate Adequate
1 1 1 1	Adequate Adequate
1 1 1 1	Adequate Adequate
1 1 1 1	Adequate
1 1 1	
1 1	Adequate
1	
	Adequate
4	Adequate
1	Adequate
1	Adequate
	er ver der und eine ver Schrift ()
3	Adequate
1	Adequate
2	Adequate
	• 27. 3).
1	Adequate
	A CONTRACTOR OF THE PARTY OF TH
	Adequate
	Adequate
1	Adequate
-	Adequate
	Adequate
	Adequate
	Adequate
1	Adequate
1	Adequate
	3 1 2

Name	Specification	Qua	ntity	Available Quantity	Adequate /Inadequate	Remar k																						
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and motorized suction, micromotor attachment with contra angle handpiece, airoter attachment, ultrasonic scaler (Piezo) with detachable autoclavable handpiece	One chair and unit per post-graduate student and Two chairs with unit for the faculty		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs with unit for the		unit per post- graduate student and Two chairs wit unit for th				
		1 Unit	2 Units	1 Unit	Adequate																							
Auto clave (fully automatic) front loading		1	1	1	Adequate																							
Steel bin		4	4	4	Adequate																							
Airoter hand pieces		2	2	2	Adequate																							
UV chamber		1	1	1	Adequate	1																						
Formalin chamber	0.000	1	1	1	Adequate	8																						
W.H.O probe		2	2	2	Adequate																							
Nabers probe		2	2	2	Adequate	1																						
Williams probe		2	2	2	Adequate																							
UNC-15 probe		4	4	4	Adequate																							
Gold Man fox probe		1	1		Adequate																							
Pressure sensitive probe		1	1	1	Adequate																							
Marquis color coded probe		1	1	10010-11	Adequate																							
Supra gingival scalers	set	2	2		Adequate																							
Sub gingival scaler	set	2	2		Adequate																							
Arkansas sharpening stone	***	1	1	1	Adequate																							
Surgical Instruments																												
Routine surgical instrument kit (Benquis periosteal elevator, periotome)	Set																											
Surgery trolleys		6	6	6	Adequate																							
X ray viewer	300015 d section 5006	1	1	1	Adequate																							
Surgical cassette with sterilisation pouches	eners an	4	4	4	Adequate																							
Electro surgery unit		1	1	1	Adequate																							
Special Surgical Inst	ruments																											
Kirkland's knife	set	1	1	1	Adequate																							
Orban's knife	set	1	1	1	Adequate																							
Paquette blade handle		1	1	1	Adequate																							
Krane kaplan pocket marker	set	1	1	1	Adequate																							
Mc Calls universal curettes	set	1	1	1	Adequate																							
Gracey's curettes (No.1-18)	set	2	2	2	Adequate																							
Mini five curettes	set	1	1	1	Adequate																							
Cumine scalar		1	1	1	Adequate																							
Mallet		1	1	1	Adequate																							
Chisel		1	1	1	Adequate	3 44 5																						
Oschenbein chisel	straight, curved	1	1	1	Adequate																							
Schluger bone file	2 Minutes 40 M	1	1	1	Adequate																							
Bone fixation screw kit		1	1	1	Adequate																							
Bone scrapper		1	1	1	Adequate																							



Bone trephines for harvesting autografts	1 set	-		1	Adequate	
Bone regenerative materials	Bone graft and GTR				Adequate	
Some regenerative materials	membranes	5	5	5		
Local drug delivery systems	At least two different agents	1 each	1 each	1 each	Adequate	
Root conditioning agent	At least two different agents	2	2	2	Adequate	Marie Constant Constant
Micro needle holder		1	1	1	Adequate	
Micro scissors		1	1	1	Adequate	
Magnifying loop (2.5 – 3.5)		1	1	1	Adequate	
Operating microscope	optional	1	1	1	Adequate	
3rd generation digital probe	optional	1	1	1	Adequate	11
Bone expander and bone crester	optional	1	1	1	Adequate	
Distraction osteogenesis kit	optional	1	1	1	Adequate	
Bone mill	optional	1	1	1	Adequate	
Bone graft / membrane placement spoon	optional	1	1	1	Adequate	
Bone condenser		1	1	1	Adequate	
Peizo-surgery unit	optional	1	1	<u>_</u>	Adequate	
Centrifuge for PRP/PRF preparation	optional	1	1	1	Adequate	
Soft tissue laser (8 watt)	Ориона	1	1		Adequate	
Osteotome	set optional	1	1	1	Adequate	*****
Miscellaneous Instrun	The state of the s				Yes	
Composite gun with material kit	leits	1	1	1	Adequate	
Splinting kit with material		2	2	2	Adequate	
Composite finishing kit		1	1	1	Adequate	
Glass Ionomer cement				-		
		1	1	11	Adequate	
Digital camera Intra Oral camera		1	1	1	Adequate	
		1	1	11	Adequate	
Ultrasonic cleaner		1	1	11	Adequate	-
Emergency kit		1	1	1	Adequate	n=== ,,,
Refrigerator		1	1	1	Adequate	
X-ray viewer		2	2	2	Adequate	
LCD projector		1	1	11	Adequate	
Computer with internet		1	1	1	Adequate	
connection with attached printer and scanner		6.0000 1771				
Implant Equipment	- CANADA	-			Adequate	
Electrical dental chair and unit	West of the second	1	1	1	Adequate	
Physio dispenser		1	1	1	Adequate	******
Implant kit	At least two different systems	2	2	2	Adequate	
Implants		10	10	10	Adequate	
Implant maintenance kit (plastic instruments)		1 set	1 set	1 set	Adequate	
Implant guide	100 Marian III III II	1	1	1	Adequate	
X-ray viewer		1	1	1	Adequate	
Needle destroyer		1	1	1	Adequate	
Ultrasonic cleaner capacity 3.5 lts		1	1	1	Adequate	and the second
Autoclave programmable for all		1	1	1	Adequate	
recommended cycles						
RVG with x-ray machine		1	1	1	Adequate	
Refrigerator		1	1	1	Adequate	
Surgical kit		2	2	2	Adequate	
Sinus lift kit		1	. 1	1	Adequate	
Educating models	7	1	1	1	Adequate	
Implant removing kit	DEENDARY	1	1	1	Adequate	

Department: Conservative Dent		Quantity	/ Available		0
Name	Specification		Quantity	Adequat e /Inadequ ate	Remark
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and motorized suction, micromotor, airotor attachment with hand pieces	unit per pos graduate student two chairs wi unit for faculty	and ith the		
T 2 2228		1 Unit 2 U	Inits 1 Unit	Adequate	
ENDOSONIC HANDPIECES – Micro endosonic Tips, retro treatment		2	2	Adequate	
Mechanised rotary instruments including hand pieces (speed and torque control) and hand instruments various systems	3	3	3	Adequate	
Rubber dam kit		1 per chair	1 per chair	Adequate	
Autoclaves for bulk instrument sterilization vacuum (Front loading)		2	2	Adequate	
Autoclaves for hand piece sterilization		1	1	Adequate	
Apex locators one for every two chairs		2	2	Adequate	
Pulp tester	V. FOOD SWITTERS (C.	2	2	Adequate	
Equipments for injectable thermoplasticized gutta percha		1	1	Adequate	
Operating microscopes 3 step or 5 step magnification		1	1	Adequate	
Surgical endo kits (Microsurgery)		2	2	Adequate	
Set of hand instruments (specifications required)		1	1	Adequate	
Sterilizer trays for autoclave		4	4	Adequate	
Ultrasonic cleaner capacity 3.5 lts		1	1	Adequate	
Variable Intensity polymerization equipments - VLC units	Desirable	1	1	Adequate	
Conventional VLC units one for every two chairs		2	2	Adequate	
Needle destroyer		2	2	Adequate	
Magnifying loupes one for students and one for faculty		1	1	Adequate	
LCD projector		1	1	Adequate	
Composite kits with different shades and polishing kits		2	2	Adequate	
Ceramic finishing kits, metal finishing kits	In ceramic labs	2	2	Adequate	
Amalgam finishing kits		2	2	Adequate	
RVG with x-ray machine developing kit		1	1	Adequate	
Chair side micro abrasion		1	1	Adequate	
Bleaching unit	i i	1	1	Adequate	
Instrument retrieval kits		1	1	Adequate	
Computer with internet connection with attached printer and scanner	DEENDAYA	1	1	Adequate	
	(all all all all all all all all all all	1	1	Adequate	

Refrigerator	1	1	Adequat e
Equipments for casting procedures		1	Adequat e
Equipments for ceramics including induction casting machines/ burnout preheat furnaces/ wax elimination furnaces	1	1	Adequat e
Lab micro motor/ metal grinders / sand blasters/ polishing lathes/ duplicator equipment/ vacuuminvestment equipments	1	1	Adequat e
Laser (preferably hard tissue)	1	1	Adequat e
Face bow with semi adjustable articulator	1	1 s	Adequat e

Department: Orthodontics and Dentofacial Orthopedics

Name				Availa ble Quan tity	Adequate /Inadequate	Rem ark
Dental Chairs and Units	Electrically operated with shadow less lamp, spittoon, 3 way syringe, instrument tray and motorized suction	PG stude Two chai for the fa	irs with unit culty			
		1 Unit	2 Units	1 Unit	Adequate	
Vacuum /pressure moulding unit		1		1	Adequate	
Hydrogen soldering unit		1		1	Adequate	
Lab micromotor		3		3	Adequate	
Spot welders		3		3	Adequate	
Model trimmer (Double disc)		2		2	Adequate	
Light curing unit		2		2	Adequate	
High intensity light curing unit		1		1	Adequate	
Polishing lathes	188	2		2	Adequate	
Tracing tables		3		3	Adequate	
SLR digital camera		1		1	Adequate	
Scanner with transparency adapter		1		1	Adequate	
X-ray viewer		3		3	Adequate	
LCD projector		1		1	Adequate	
Autoclaves for bulk instrument Sterilization vacuum (Front loading)		1		1	Adequate	
Needle destroyer		1		1	Adequate	
Dry heat sterilizer		1	-	1	Adequate	
Ultrasonic scaler		1		1	Adequate	51102
Sets of Orthodontic pliers		3		3	Adequate	
Orthodontic impression trays		3		3	Adequate	
Ultrasonic cleaner capacity 3.5 Its		1		1	Adequate	
Electropolisher		1		1	Adequate	
Typodonts with full teeth set		3		3	Adequate	
Anatomical articulator with face bow attachments		1			Adequate	
Free plane articulators		1	-1	1	Adequate	
Hinge articulators		4		4	Adequate	
Computer software for cephalometrics		1		1	Adequate	
Computer with internet connection with attached printer and scanner	DEENDAYALU	1		1	Adequate	
Refrigerator	The state of the s	1		1	Adequate	

Name	Specification	Qua	ntity	Available Quantity	Adequate /Inadequate	Remark
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument	unit pe	r post- :e			45
	tray and suction	one cha unit for Faculty	SECOND THE WASHING			
		1 Unit	2 Units	1 Unit	Adequate	
RVG with intra oral radiography machine (FDA Approved)	55-70 Kvp with digital Compatibility	1		1	Adequate	
Extra oral radiography machine	100 kvp	1		1	Adequate	
Panoramic radiography (OPG) machine with cephalometric and TMJ attachment with printer	The contract of the contract o	1		1	Adequate	
Intra-oral camera		1		1	Adequate	· · · · · · · · · · · · · · · · · · ·
Pulp tester		2		2	Adequate	
Autoclave		1		1	Adequate	
Punch biopsy tool		2		2	Adequate	
Biopsy equipment		1	3 22 7	1	Adequate	
Surgical trolley	7 2	2		2	Adequate	
Emergency medicines kit		1		1	Adequate	
Extra oral cassettes with intensifying screens (Conventional and rare earth)		4		4	Adequate	
Lead screens		2		2	Adequate	
Lead aprons		2		2	Adequate	
Lead gloves		2		2	Adequate	
Radiographic filters (Conventional and rare earth)		1		1	Adequate	, vse.
Dark room with safe light facility		1		1	Adequate	
Automatic radiographic film processors		2		2	Adequate	
Radiographic film storage lead containers		1		1	Adequate	
Thyroid collars		1		1	Adequate	
Digital sphygmomanometer		1		1	Adequate	
Digital blood glucose tester		1		1	Adequate	
Digital camera		1		1	Adequate	
X-ray viewer boxes		2		2	Adequate	
Lacrimal probes		2 sets	100000	2 sets	Adequate	
Sialography cannula		2 sets	e v	2 sets	Adequate	
Illuminated mouth mirror and probe		2		2	Adequate	
Computer with internet connection with attached printer and scanner		1		1	Adequate	
LCD projector		1		1	Adequate	
Refrigerator		1	-	1	Adequate	



Dean / Principal Stamp & Signature

Dr. Birongone R.S

PRINCIPAL

Pt. Deenda Val Upadhyay

Dental College, Solapur