EQUIPMENT REQUIREMENTS

As per Revised BDS Course Regulations, 2007

Department: Oral Medicine and Radiology

			antity	Available	0 -1	D1
Name	Specification	50 Intake	100 Intake	Quantity	Adequate /Inadequa te	Remark
	(Space allocation for each dental chair 100 Sq.Ft. in All the Departments)					
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventry Suction, X-Ray viewer, 3 way syringe, instrument tray, Dental Operator's Stool with height adjustment		12	12	Adequate	
Panoramic X-Ray with Cephalometric	Preferably digital		1	1	Adequate	
Intra Oral X-ray Unit	70 KV, 8 mA, high frequency preferably Digital timer		3	3	Adequate	- 1
Pulp testers	Digital				Adequate	
Automatic periapicalX- ray Developer	· · · · · · · · · · · · · · · · · · ·		1	1	Adequate	
Automatic Panoramic with Cephalometric X-ray Developer			1	No		
X-ray viewer	For Panoramic and Cephalometric films		2	2	Adequate	
Radiovisiography	Digital Intra X-ray system with one Sensor andSoftware		1	1	Adequate	-
GeneralX-ray Unit			1	1	Adequate	
Ortho Pantmograph	Preferably digital		1	1	Adequate	
Automatic Developers / Dark Room			2	2	Adequate	
Lead aprons			2 .	2	Adequate	
Lead Gloves			1	1	Adequate	
X-ray Hangers			6	6	Adequate	
X-ray Viewers	2 2		2	2	Adequate	
Diagnostic Kits	Mouth mirror, dental probe, college tweezers		40	40	Adequate	
Lead Screen			1	1	Adequate	
Biopsy Kit			1	1	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		2	2	Adequate	
Computers	Minimum Pentium – IV		1	1	Adequate	
Intra Oral Camera	With High resolution		1	1	Adequate	
Needle Burner with Syringe Cutter			2	2	Adequate	

Department: Oral Pathology and Oral Microbiology

	3,					
	Specification	Quantity		Available	Adequate	Remark
Name	Specification	50	100	Quantity		Remark
		Intake	Intake		/Inadequat	
	*			9	е	
Dental Chairs and	With shadowless lamp, spittoon, 3 way syringe,	-	2	2	Adequate	×
Units	instrument tray and suction					
Microscopes			40	. 40	Adequate	
Microtome			1	1	Adequate	
Wax bath			1	1	Adequate	
Water bath			1	1	Adequate	
Knife sharpener			1	1	Adequate	
Hot plate			1	1	Adequate	
Spencer Knife			1	1	Adequate	

Department: Public Health Dentistry

		Quantity		Available	Adoquato	Damanlı
Name	Specification	50 Intake	100 Intake	Quantity	Adequate /Inadequat e	Remark
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, micromotor control light cure 3 way syringe, X-ray viewer, instrument tray Dental Operators stool with height adjustment with shadowless lamp, spittoon, 3 way syringe, instrument tray and suction,		16	16	Adequate	
Autoclaves	micromotor, airotor, light cure Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		2	2	Adequate	
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket with digital timer		2	2	Adequate	
Needle burner with syringe cutter			4	4	Adequate	
MOBILE CLINIC	š				Adequate	
Mobile dental van	Mobile dental van with two dental chairs with all the attachments and adequate sitting space for 15 to 20 people				Adequate	,
Dental chair with unit	Hydraulically operated with spittoon attachment, halogen light with 2 intensity, air ventury suction, airrotor, micromotor, 3 way-scaler and light cure, x-ray viewer, instrument tray, operating stool	1			Adequate	
Autoclaves	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	-	1	1	Adequate	
Intraoral x-ray	Portable, 70 KV, 8 mA		1	1	Adequate	
Glass bead sterilisers			1	1	Adequate	
Compressor	1.25 HP		1	1	Adequate	
Metal Cabinet	With wash basin		1	1	Adequate	
Portabledental chair	Suitcase unit with airotor, micromotor, scaler and compressor 0.25 HP		2	2	Adequate	
Stabiliser	4 KV		1	1	Adequate	
Generator	4 KV		1	1	Adequate	
Water tank	400 ltrs		1	1	1	
Oxygen cylinder			1	Yes	1	8.

Publicaddress		1	Yes	Adequate	
system	= 0				*
TV and video		1	1	Adequate	
cassette player					
Demonstration		1	1	Adequate	6
models					

Department: Paediatric and Preventive Dentistry

Departmen	nt: Paediatric and Preventive Dentistry	0	_ 4:4	A ! - -		I
Name	Specification		ntity	Available	Adequat	Remark
rume	Specification	50	100	Quantity	е	
r		Intake	Intake		/Inadequ	
					ate	
Dental Chairs and	Electrically operated, Spittoon attachment, Halogen		20 .	20	Adequate	
Units	Light with 2 intensity, high power motorized	1				
	evacuation system, Air rotor with miniature, Airrotor	i				
	HPS, Micro motor with miniature contrangle Hand					
2	piece, 3 way syringe, ultrasonic sealer with 3 tips,	ł				
	Light cure unit LED based heat free, X-ray viewer,	1				
	instrument tray Dental Operator's stool with height					
A	adjustment (Pedo chair preferred)				0.1	
Autoclaves	Having wet and dry cycle, which can achieve 135°C		2	2	Adequate	
	with minimum capacity of 20 liters				ļ	ļ
Ultrasonic cleaner		-	2	2	Adequate	
	digital timer					
Needle Burner	1		3	3	Adequate	
With syringe	¥ . *					
cutter	*					
Amalgamator		8	1	1	Adequate	
Pulp Tester Digital			1	1	Adequate	
Rubber dam kit			5	5	Adequate	
for pedo						
Apex locator			1	1	Adequate	
Endo motor	With torque control HPs		1	1	Adequate	
Injectable gutta			1	1	Adequate	
percha with						
condensation						-
Radiovisiography	Digital intra X-ray system with pedo sensor and		1	1	Adequate	
, and the graph of	software		-	-		10.
Intra Oral Camera	With high resolution		1	1	Adequate	
Scale instruments	Vittingiresolution		10	10	Adequate	-
Restorative		-	10	10	Adequate	-
instruments			10	10	Auequate	
			10	10	Adamusta	
Extraction forceps		-	10	10	Adequate	
Intra-oral X-ray		-	1	1	Adequate	-
Automatic			1	1	Adequate	
Developer		-			1	
Computer	Minimum Pentium – IV		1	1	Adequate	
	PEDO LAB					
Plaster dispenser	One each for plaster and stone plaster		2	2	Adequate	
Model Trimmer	With diamond disc		1	1	Adequate	
Model Trimmer	Double disc one diamond and one carborandam disc		1	1	Adequate	
Welder with			1	1	Adequate	
soldering	9 9					
attachments						
Vibrator			2	2	Adequate	
	I .				/58	

Lab micro motor	Heavy duty with Hps	3	3	Adequate	
Dental Lathe		1	1	Adequate	
Model Trimmer		1	1	Adequate	160
Steam cleaner		1	1	Adequate	

Steam cleaner		1	1	Adequate	
Pressure		1	1	Adequate	
moulding	r e		,		
Machine			9		
Carborandum		1	1	Adequate	
Disc					
Diamond disc		1	1	Adequate	

Department: Orthodontics

		Qua	ntity	Available	A 1	D
Name	Specification	50	100	Quantity	Adequat	Remark
		Intake	Intake		e /Inadequ ate	2
Dental Chairs and	Electrically operated, Spittoon attachment, Halogen		18	18	Adequate	
Units	Light with 2 intensity, Air Ventury suction, micromotor control light cure 3 way syringe, X-ray viewer, instrument tray Dental Operators stool with height adjustment		- 1			
Unit mount scaler			5	5	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		2	2	Adequate	
	Minimum capacity 13 litres with mesh bucket with digital timer	* " =	2	2	Adequate	
ORTHO LAB					Adequate	
Plaster dispenser	One each for plaster and stone plaster		2	2	Adequate	
Vibrator			2	2	Adequate	
Model Trimmer			2 .	2	Adequate	
Micromotor	Heavy duey		4	4	Adequate	
Lathe			2	2	Adequate	
X-ray viewers			2	2	Adequate	
OPG with cephalostat	If available in radiology its is adequate		1	1	Adequate	
Welders		,	4	4	Adequate	
Blue Torch	*		1	1	Adequate	
Base Formers			4	4	Adequate	
Typodont			4	4	Adequate	
Set of pliers			10	10	Adequate	
Welder with			1	1	Adequate	n 22
soldering						
attachments						* 2
Hydro solder			1	1	Adequate	
Typhodont	With metal teeth wax rim of Class I, II, III		4	4	Adequate	
articulator						
Pressure			1	1		
moulding						
machine						

Department: Periodontology

	Specification	Quantity		Available	Adequat	Remark
Name		50 Intake	100 Intake	Quantity	e /Inadequ	Remark
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, with shadowless lamp, spittoon, 3 way syringe, Ultrasonic sealer with 3 tips, X-ray viewer, instrument tray Dental Operators stool with height adjustment	×y	34	34	Adequate	
Scaling instrument sets			10	10	Adequate	
Surgical instrument sets			6	6	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	*	3	3	Adequate	,
Ultrasonic scaler	Minimum capacity 13 liters with mesh bucket		2	2	Adequate	
Electrosurgical Cautery			1	1	Adequate	
Needle burner With syringe cutter			6	6	Adequate	
LASER	Soft tissue laser		1	1	Adequate	
Surgical motor With physio dispenser			1	1	Adequate	

Department: Conservative Dentistry and Endodontics

		Quantity		Available	^ -l	Remark
Name		50 Intake	100 Intake	Quantity	Adequat e /Inadequ ate	Remark
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, with shadowless lamp, spittoon, 3 way syringe, Ultrasonic sealer with 3 tips, X-ray viewer, instrument tray Dental Operators stool with height adjustment		34	34	Adequate	
Rubber dam kits			6	6	Adequate	
Restorative instruments kits		9	10	10	Adequate	
R.C.T. instrument kits	. 48	15	10	10	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		3	3	Adequate	
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket		2	2	Adequate	
NeedleBurner With Syringe Cutter		9	4	4	Adequate	
Amalgamator	With auto proportion, auto dispenser		3	3	Adequate	
Rubber dam kits			6	6		

Pulp Tester Digital		4	4	Adequate	
Apex locator		2	2	Adequate	
Glass bead		6	6	Adequate	2
sterilizers				Adequate	
Plaster dispenser		2	2	Adequate	
Vibrator		2	2	Adequate	
Ceramic Unit		1	1	Adequate	
Casting machine		1	1	Adequate	
Intra-oral X-ray Unit	Proper radiation safety	1	1	Adequate	
Automatic		1	1	Adequate	
Developer	, ,				
Radiovisiography	RVG with Computer	1	1	Adequate	
Endo motor	With torque control Hps	1	1	Adequate	
Bleaching Unit		1	1	Adequate	
Magnification loops		2	2	Adequate	**
Injectabl gutta percha		2	2	Adequate	
PHANTOM LAB UNIT	Phantom Table fitted with Halogen Operating Light Phantom Head body type neck joint for all the movement, TMJ movement, Modular with Air rotor, Micro motor with contra angle Hps, 3 way syringe, jaw with ivorine teeth, preferably soft gingival, dental operator's stool (not to use extracted or cadaver teeth).	60	60	Adequate	. • .

Chemical Laboratory

		Qua	ntity	Available		Remark
Name	Specification	50	100	Quantity	Adequat e	
		Intake	Intake	,	/Inadequ	140
Plaster Dispenser	One each for plaster and stone plaster		2	2	Adequate	
Model Trimmer	Carborandum disc		1	1	Adequate	
	Diamond disc		1	1	Adequate	
Lathe	Heavy Duty		2	2	Adequate	
Lab Micromotor	With heavy duty handpiece		3	3	Adequate	
Ultrasonic cleaner	Minimum capacity 5 litres		1	1	Adequate	
Spindle Grinder			1	1	Adequate	
Vibrator			2	2	Adequate	
Burnout furnace	,		1	1	Adequate	W.
Porcelain furnace			1	1	Adequate	
Sandblasting Machine			1	1	Adequate	
Lab Airotor			1	1	Adequate	
Pindex System			1	1	Adequate	
Circular saw	4		1	1	Adequate	4
Vacuum mixer			1	1	Adequate	
Pnematic chisel			1	1	Adequate	
Casting machine	Motor cast / induction casting preferred		1	1		

Department: Oral & Maxillofacial Surgery

A) Exodontia

		Qua	ntity	Available	Adamusta	Domorle
Name	Specification	50 Intake	100 Intake	Quantity	Adequate /Inadequat e	Remark
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, X-ray viewer, 3 way syringe, instrument tray Dental Operators stool and height adjustment and suction, Micromotor / Airotor		30	30	Adequate	
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters		2	2	Adequate	
Ultrasonic cleaner	Minimum capacity 13 litres with mesh bucket		2	2	Adequate	
Needle Burner with Syringe Cutter			6	6	Adequate	
Extraction forceps sets	Complete set		20	20	Adequate	
Dental elevators	Complete set		10	10	Adequate	
Minor Oral Surgery kits			6	6	Adequate	
Emergency drugs tray			1	1	Adequate	1
Oxygen cylinder with mask			1	1	Adequate	
X-ray viewers			2	2	Adequate	
Computer	Minimum Pentium – IV		1	Yes		

B) Minor Surgery

		Qua	ntity	Available	Adequ	Remark
Name	Specification	50	100	Quantity	ate	Kelliaik
		Intake	Intake		/Inadeq	
		3			uate	=
Dental Chairs and	Electrically operated, Spittoon attachment, Halogen Light		5	5	Adequate	
Units	with 2 intensity, high power evaculation system, Air ventury					
	suction, X-ray viewer, Airrotor, Micromotor with straight	1				· ir
	and contra-angle Handpiece, 3 way syringe, instrument tray Dental					
	Operators stool and height adjustment and suction.	ES A		2		н

Department: Prosthodontics and Crown & Bridge

Departmen	it: Prosthodontics and Crown & Bridge					
		Qua	ntity	Available	Adequat	Remark
Name	Specification	50	100	Quantity		Keillaik
		Intake	Intake		e /Inadequ	
					ate	
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, 3 way syringe, Ultrasonic sealer with 3 tips, Light cure unit, instrument tray and suction, Dental Operators stool with height adjustment		34	34	Adequate	
Semi-adjustable articulator	With face bow		2	2	Adequate	

Extra oral / intra		2	2	Adequate	
oral tracer		2	2	Aucquate	
Dewaxing unit		2	2	Adequate	
Curing unit	*	2	2	Adequate	
Dentalcasting				Adequate	-
machine		1 1	1	Adequate	
					* 2
Wax burnout		1	1	Adequate	3
Furnace					
Preheating		1 1	1	Adequate	
furnace					
Surveying unit		2	2	Adequate	
Heavy duty hand	Lab micromotors	4	4	Adequate	
Piece					
Autoclave	Having wet and dry cycle, which can achieve 135°C	2	2	Adequate	et .
	with minimum capacity of 20 liters				
NeedleBurner	*	2	2	Adequate	
With Syringe				- a	
Cutter					
Plaster dispenser	One each for plaster and stone plaster	2	2	Adequate	
ModelTrimmer		1	1	Adequate	
with					
carborandum					
Disc				-	
Model Trimmer		2	2	Adequate	£
with diamond					
Disc	,				
Acryliser		3	3	Adequate	
Lathe		2	2	Adequate	
Flask press		4	4	Adequate	
Deflasking unit		4	4	Adequate	
Dewaxing unit		3	3	Adequate	
Hydraulic Press		3	3	Adequate	
Mechanical Press		2	2	Adequate	
Vacuum mixing		1	1	Adequate	
machine		1	_	lacquate	
Lab Micro motor	With heavy duty handpiece	4	4	Adequate	*
Curing pressure	Trial neary daty nanapiece	1	1	Adequate	
Pot		-		racquate	
Porcelain furnace		2	2	Adequate	
Vibrator		2	2	Adequate	
Sandblasting unit			2	Adequate	
Ultrasonic cleaner		2		Adequate	
Model Trimmer		2	2		
		4	4	Adequate	
Hot water sterilizer		2	2	Adequate	
	Commonwed hoth			A -l	
Geyser	Compound bath	2	2	Adequate	
H.P. grinder with		3	3	Adequate	
suction				1	
Heavy duty lathe		2	2	Adequate	1
Phantom heads		50	50	Adequate	
Pre-clinical	Gas connection & Bunsen burner	100	100	Adequate	
working tables					

Ceramic and Cast Partial Laboratory

that the second		Qua	ntity	Available		D
Name	Specification	50	100	Quantity	Adequat	Remark
	2"	Intake	Intake		e	
					/Inadequ	
81 !					ate	
Plaster dispenser	One each for plaster and stone plaster		2	2	Adequate	
Duplicator			1	1	Adequate	
Pindex System			1	1	Adequate	
Circular saw	4		1	1	Adequate	
Burnout furnace			1	1	Adequate	
Sandblasting Machine	With two containers		1	1	Adequate	
Electro polisher	/		1	1	Adequate	
Model Trimmer with Carborandum disc			1	1	Adequate	
Model Trimmer with diamond disc			1	1	Adequate	
Induction casting machine			1	1	Adequate	
Programmable porcelain furnace with			1	1	Adequate	
vacuum pump with instrument kit and			-	_		
material kit		1			-	
Spot welder with soldering attachment			1	1	Adequate	
of cable	*,			_		
Vacuum mixing machine			1	1	Adequate	
Steam Cleaner			1	1	Adequate	
Spindle Grinder 24000 rpm with			1	1	Adequate	
vacuum suction			_	_		
Wax heater			1	1	Adequate	
Wax carver			1	1	Adequate	
Curing pressure pot			1	1	Adequate	1.
Milling machine			1	1	Adequate	1
Heavy duty lathe with suction			1	1	Adequate	
Preheating furnace			1	1	Adequate	-
Palatal trimmer			1	1	Adequate	-
Ultrasonic cleaner	5 liters capacity		1	1	Adequate	
Composite curing unit	3 iters capacity		1	1	Adequate	-
Micro surveyor		-	1	1	Adequate	
Pre-clinical Prosthetic Laboratory	Work table preferably				Adequate	
Fre-clinical Frostnetic Laboratory	Work table preferably complete stainless steel fitted		60	60	Auequate	
	with light, Bunsen burner, air	1	20	20	100	
	blower, working stool.		20	20		
	Adequate number of lab micro			v.		
	motor with attached hand		100			
	piece.					
Plaster room for pre-clinical work	p.ccc.	-	-		Adequate	1
Plaster dispenser	One each for plaster and stone	-	2	2	Adequate	-
Vibrator	one each for plaster and storie		2	2	Adequate	1
Lathe		1	2	2	Adequate	+
Model Trimmer		-	1	1	Adequate	
Carborandum Disc		-	1		Adequate	-
Diamond disc				1		-
Diamona disc		-	1	1	Adequate	
		-	1	-		
		-	1	-		
	L					

EQUIPMENT REQUIREMENTS

Name	Specification	Quantity	Availab le Quantit y	Adequate /Inadequat e	Remark
Electrical Dental Chairs and Units	motorized suction,				
		1 Unit 2 Units	Yes	Adequate	-
Articulators – semi adjustable/ adjustable		6		Adequate	
with face bow	4			Adequate	
Micromotor – (Lab Type can also be attached		2	2	Adequate	
fixed) to wall		2		Adequate	
Jitrasonic scaler		2	2	Adequate	
ight cures		2	2	Adequate	
Hot air oven		1	1	Adequate	
Autoclave	· · · · · · · · · · · · · · · · · · ·	2	2	Adequate	
Surveyor		2	2	Adequate	
Refrigerator		1	1	Adequate	-
X-ray viewer		1	1	Adequate	-
Pneumatic, Crown bridge remover		2	2	Adequate	
Veedle destroyer		1	1	Adequate	
Intra oral camera		1	1	Adequate	
Digital SLR camera		1	1	Adequate	
Computer with internet connection with		1	1	Adequate	-
attached printer and scanner		1	1	Aucquate	
LCD projector		1	1		
Clinical Lab for Prosthet	ics		-	-	1
Plaster dispenser		2	2	Adequate	+
Model trimmer with carborandum Disc		1	1	Adequate	
Model trimmer with diamond disc		1	1	Adequate	
High speed lathe		2	2	Adequate	
Vibrator		2	2	Adequate	
Acrylizer		1	1	Adequate	
Dewaxing unit		1	1	Adequate	
Hydraulic press		1	Yes	Adequate	
Mechanical press		1	1	Adequate	
Vacuum mixing machine		1	1	Adequate	1
Micro motor lab type		2	2	Adequate	
Curing pressure pot		1	1	Adequate	
Pressure molding machine		1	1	Adequate	1
Chrome – Cobalt Lab Equip	ment	_	 		1
Duplicator	<u> </u>	1	1	Adequate	1
Pindex system		1	1	Adequate	1
Burn-out furnace		1	1	Adequate	1
Welder		1	1	Adequate	-
Sandblaster	Micro and macro	1	1	Adequate	1
Electro – polisher		1	1	Adequate	+
Model trimmer with carborandum disc		1	1	Adequate	

Model trimmer with diamond disc			1	Adequate
Model trimmer with double disc (one		1		Adequate
Carborandum and one				
diamond disc)			* -	
Casting machine, motor cast with		1	1	Adequate
the safety door closure, gas blow torch with		2		
regulator	*			
Dewaxing furnace		1	1	Adequate
Induction casting machine with vacuum pump,		1	1	Adequate
capable of casting				E = 2
titanium chrome cobalt precision metal				
Spot welder with soldering, attachment of		1	1	Adequate
cable		,		
Steam cleaner		1	1	Adequate
Vacuum mixing machine		1	1	Adequate
Spindle grinder 24,000 ROM with vacuum		1	1	Adequate
suction		2 1		
Wax heater		2	2	Adequate
Wax carvers (Full PKT Set)		2	2	Adequate
Milling machine		1	1	Adequate
Stereo microscope		1	1	Adequate
Magnifying work lamp	E	1	1	Adequate
Heavy duty lathe with suction	*	1	1	Adequate
Preheating furnace	,	1	1	Adequate
Dry model trimmer		1	1	Adequate
Die cutting machine		1	1	Adequate
Ultrasonic cleaner		1	1	Adequate
Composite curing unit	8	1	1	Adequate
Ceramic Lab Equipment				
Fully programmable porcelain furnace with		1	1	Adequate
vacuum pump				
Ceramic kit (instruments)		3	3	Adequate
Ceramic materials (kit)		1	1	Adequate
Ceramic polishing kit		2	2	Adequate
Implan E quipment				
Electrical dental chair and unit	1 12	1	1	Adequate
Physio dispenser		1	1	Adequate
Implant kit	Minimum 2 systems	2	2	Adequate
Implants		10	10	Adequate
Prosthetic components		10	10	Adequate
Unit mount light cure		1	1	Adequate
X-ray viewer		1	1	Adequate
Needle destroyer		1	1	Adequate
Ultrasonic cleaner capacity 3.5 lts		1	1	Adequate
Autoclave programmable for all recommended cycles	·	1	1	Adequate
X-ray machine with RVG		1	1	Adequate
Refrigerator		1	1	Adequate
Surgical kit/prosthetic kit		2	2	Adequate
Educating models		1	1	Adequate
Implant removing instruments		1	1	Adequate
			1 -	

Department: Periodontology						,
Name	Specification	Qua	ntity	Available Quantity	Adequate /Inadequate	Remar k
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and motorized suction, micromotor attachment with contra angle handpiece, airoter attachment, ultrasonic scaler (Piezo) with detachable autoclavable handpiece	One chair and unit per post-graduate student and Two chairs with unit for the faculty				
	f.,	1 Unit	2 Units	1 Unit	Adequate	
Auto clave (fully automatic) front loading		1	1	1	Adequate	
Steel bin		4	4	4	Adequate	
Airoter hand pieces		2	2	2	Adequate	
UV chamber		1	1	1	Adequate	
Formalin chamber	. " 4	1	1	1	Adequate	
W.H.O probe		2	2	2	Adequate	
Nabers probe		2	2	2	Adequate	
Williams probe		2	2	2	Adequate	
UNC-15 probe		4	4	4	Adequate	
Gold Man fox probe		1	1	1	Adequate	
Pressure sensitive probe		1	1	1	Adequate	
Marquis color coded probe		1	1	1	Adequate	
Supra gingival scalers	set	2	2	2	Adequate	
Sub gingival scaler	set	2	2	2	Adequate	
Arkansas sharpening stone		1	1	1	Adequate	
Surgical Instruments	-					
Routine surgical instrument kit (Benquis periosteal elevator, periotome)	Set			3		
Surgery trolleys		6	6	6	Adequate	
X ray viewer		1	1	1	Adequate	
Surgical cassette with sterilisation pouches		4	4	4	Adequate	
Electro surgery unit		1	1	1	Adequate	
Special Surgical Inst	ruments					
Kirkland's knife	set	1	1	1	Adequate	
Orban's knife	set	1	1	. 1	Adequate	
Paquette blade handle		1	1	1	Adequate	
Krane kaplan pocket marker	set	1	1	1	Adequate	
Mc Calls universal curettes	set	1	1	1	Adequate	
Gracey's curettes (No.1-18)	set	2	2	2	Adequate	
Mini five curettes	set	1	1	1	Adequate	
Cumine scalar		1	1	1	Adequate	
Mallet		1	1	1	Adequate	
Chisel		1 .	1	1	Adequate	
Oschenbein chisel	straight, curved	1	1	1	Adequate	
Schluger bone file		1	1	1	Adequate	
Bone fixation screw kit		1	1	1	Adequate	
Bone scrapper	-	1	1	1	Adequate	

Bone trephines for harvesting autografts	1 set	T			1	Adequate
Bone regenerative materials	Bone graft and G	TR				Adequate
	membranes	1	5	5	5	
Local drug delivery systems	At least two differe				1 each	Adequate
9	agents		cucii	1 cacii	I cacii	
Root conditioning agent	At least two differe	ent	2	2	2	Adequate
	agents		_		_	
Micro needle holder			1	1	1	Adequate
Micro scissors			1	1	1	Adequate
Magnifying loop (2.5 – 3.5)	2		1	1	1	Adequate
Operating microscope	optional		1	1	1	Adequate
3rd generation digital probe	optional		1	1	1	Adequate
Bone expander and bone crester	optional		1	1	1	Adequate
Distraction osteogenesis kit	optional		1	1	1	Adequate
Bone mill	optional		1	1	1	Adequate
Bone graft / membrane placement spoon	optional		1	1	1	Adequate
Bone condenser		-	1	1	1	Adequate
Peizo-surgery unit	optional		1	1	1	Adequate
Centrifuge for PRP/PRF preparation	optional		1	1	1	Adequate
Soft tissue laser (8 watt)	Ортіона		1	1	1	Adequate
Osteotome	set optional		1	1	1	Adequate
Miscellaneous Instrun						Yes
Composite gun with material kit			1	1	1	Adequate
Splinting kit with material			2	2	2	Adequate
Composite finishing kit						
			1	1	1	Adequate
Glass Ionomer cement	2		1	1	1	Adequate
Digital camera			1	1	1	Adequate
Intra Oral camera			1	1	1	Adequate
Ultrasonic cleaner			1	1	1	Adequate
Emergency kit			1	1	1	Adequate
Refrigerator	,		1	1	1	Adequate
X-ray viewer			2	2	2	Adequate
LCD projector			1	1	1	Adequate
Computer with internet			1	1	1	Adequate
connection with attached printer and scanner						
Implant Equipment						Adequate
Electrical dental chair and unit			1	1	1	Adequate
Physio dispenser			1	1	1	Adequate
Implant kit	At least two difference	ent	2	2	2	Adequate
	systems					N. /
Implants			10	10	10	Adequate
Implant maintenance kit (plastic instruments)		1	L set	1 set	1 set	Adequate
Implant guide			1	1	1	Adequate
X-ray viewer			1	1	1	Adequate
Needle destroyer			1	1	1	Adequate
Ultrasonic cleaner capacity 3.5 lts			1	1	1	Adequate
Autoclave programmable for al			1	1	1	Adequate
recommended cycles		1,1				
RVG with x-ray machine			1	1	1	Adequate
Refrigerator			1	1	1	Adequate
Surgical kit			2	2	2	Adequate
Sinus lift kit			1	1	1	Adequate
Educating models			1	1	1	Adequate
Implant removing kit			1	1	1	Adequate

Department: Conservative Dentistry and Endodontics

Namo	Specification	Quantity		Available	Adequat	Remark
Name	Specification			Quantity	e	
					/Inadequ	
	E				ate	
Dental Chairs and Units	Electrically operated with					
	shadowless lamp, spittoon,					-
	3 way syringe, instrument tray and motorized suction,				- 1	
	micromotor,airotor	two chai	and rs with			
	attachment with	unit fo	1901 1901000000			
	hand pieces	faculty				
		1 Unit	2 Units	1 Unit	Adequate	
ENDOSONIC HANDPIECES –		2	2 Units		Adequate	
Micro endosonic Tips, retro treatment					racquate	
Mechanised rotary instruments including hand		3		3	Adequate	-
pieces (speed						
and torque control) and hand instruments			. '='			
various systems				œ.		
Rubber dam kit		1 per		1 per	Adequate	
		chair		chair		
Autoclaves for bulk instrument sterilization		2		2	Adequate	
vacuum (Front	47					
loading)	,					
Autoclaves for hand piece sterilization	4 1	1		1	Adequate	
Apex locators one for every two chairs		2		2	Adequate	
Pulp tester		2		2	Adequate	
Equipments for injectable thermoplasticized		1		1	Adequate	
gutta percha					A -1 1 -	<u> </u>
Operating microscopes 3 step or 5 step magnification		1		1	Adequate	
Surgical endo kits (Microsurgery)		2		2	Adequate	-
Set of hand instruments (specifications		1		1	Adequate	-
required)		1		1	Aucquate	
Sterilizer trays for autoclave		4		4	Adequate	1
Ultrasonic cleaner capacity 3.5 lts		<i>±</i> 1		1	Adequate	
Variable Intensity polymerization equipments	Desirable	1		1	Adequate	ļ
- VLC units	Desirable	-		†	hucquate	
Conventional VLC units one for every two		2		2	Adequate	-
chairs		_		_		
Needle destroyer		2		2	Adequate	
Magnifying loupes one for students and one		1		1	Adequate	
for faculty \	×					
LCD projector		1		1	Adequate	
Composite kits with different shades and		2		2	Adequate	
polishing kits					*	
Ceramic finishing kits, metal finishing kits	In ceramic labs	2		2	Adequate	
Amalgam finishing kits		2		2	Adequate	
RVG with x-ray machine developing kit		1		1	Adequate	
Chair side micro abrasion		1		1	Adequate	
Bleaching unit		1		1	Adequate	
Instrument retrieval kits		1		1	Adequate	
Computer with internet connection with attached printer and scanner		1		1	Adequate	
		1		1	Adequate	

Refrigerator	1	1	Adequat
Equipments for casting procedures		1	Adequat
Equipments for ceramics including induction	1	1	Adequat
casting machines/ burnout preheat furnaces/			е
wax elimination furnaces			
Lab micro motor/ metal grinders / sand	1	1	Adequat
blasters/ polishing lathes/ duplicator			e
equipment/ vacuuminvestment equipments			
Laser (preferably hard tissue)	1	1	Adequat
			e
Face bow with semi adjustable articulator	1	1s	Adequat
			e

Department: Orthodontics and Dentofacial Orthopedics

Department: Orthodontics and Der	Toraciai Orthopedics					
Name	Specification	Quantity		Availa ble Quan tity	Adequate /Inadequate	Rem ark
Dental Chairs and Units	Electrically operated with	One cha	ir & unit per			
4	shadow less lamp,	PG stude	ent and			
	spittoon, 3 way syringe,	Two cha	irs with unit			
*	instrument tray and	for the fa	aculty		,	
a contract to the contract to	motorized suction					
		1 Unit	2 Units	1 Unit	Adequate	
Vacuum /pressure moulding unit		1		1	Adequate	
Hydrogen soldering unit		1		1	Adequate	
Lab micromotor		3		3	Adequate	
Spot welders .	5 ta	3		3	Adequate	
Model trimmer (Double disc)		2		2	Adequate	
Light curing unit		2		2	Adequate	
High intensity light curing unit		1		1	Adequate	
Polishing lathes		2		2	Adequate	100
Tracing tables		3		3	Adequate	
SLR digital camera		1		1	Adequate	
Scanner with transparency adapter		1		1	Adequate	
X-ray viewer		3		3	Adequate	
LCD projector		1		. 1	Adequate	1
Autoclaves for bulk instrument Sterilization		1		1	Adequate	
vacuum (Front loading)						
Needle destroyer		1		1	Adequate	
Dry heat sterilizer		1		1	Adequate	
Ultrasonic scaler		1		1	Adequate	
Sets of Orthodontic pliers		3		3	Adequate	
Orthodontic impression trays		3		3	Adequate	
Ultrasonic cleaner capacity 3.5 lts		1		1	Adequate	
Electropolisher		1		1	Adequate	
Typodonts with full teeth set		3		3	Adequate	
Anatomical articulator with face bow		1		1	Adequate	
attachments						
Free plane articulators		1		1	Adequate	
Hinge articulators		4		4	Adequate	
Computer software for cephalometrics		1		1	Adequate	
Computer with internet connection with attached		1		1	Adequate	
printer and scanner					-	
Refrigerator		1		1	Adequate	

Department: Oral Medicine and Radiology

Name	Specification	unit per post- graduate student and one chair with unit for the		Available Quantity	Adequate /Inadequate	Remark
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and suction					
*		Faculty		411.11	A.I.	
RVG with intra oral	55-70 Kvp with digital	1 Unit	2 Units	1 Unit	Adequate Adequate	
radiography machine (FDA Approved)	Compatibility	1		1	Adequate	
Extra oral radiography machine	100 kvp	1		1	Adequate	
	Digital compatibility	1		1	Adequate	
cephalometric and TMJ attachment with printer	Digital compatibility			_	Adequate	
Intra-oral camera		1		1	Adequate	-
Pulp tester		2		2	Adequate	
Autoclave		1		1	Adequate	
Punch biopsy tool	v	2		2	Adequate	
Biopsy equipment		1		1	Adequate	
Surgical trolley		2		2	Adequate	3
Emergency medicines kit		1		1	Adequate	
Extra oral cassettes with intensifying screens (Conventional and rare earth)		4		4	Adequate	
Lead screens		2		2	Adequate	160
Lead aprons		2		2	Adequate	
Lead gloves	2.1	2		2	Adequate	
Radiographic filters (Conventional and rare earth)		1		1	Adequate	
Dark room with safe light facility		1	8 -	1	Adequate	
Automatic radiographic film processors		2		2	Adequate	
Radiographic film storage lead containers		1		1	Adequate	
Thyroid collars		1	*	1	Adequate	
Digital sphygmomanometer	V*	1		1	Adequate	
Digital blood glucose tester		1		1	Adequate	
Digital camera		1		1	Adequate	
X-ray viewer boxes		2		2	Adequate	
Lacrimal probes		2 sets		2 sets	Adequate	
Sialography cannula		2 sets		2 sets	Adequate	
Illuminated mouth mirror and probe		2		2	Adequate	
Computer with internet connection with attached printer and scanner		1	7	1	Adequate	-
LCD projector		1		1	Adequate	
Refrigerator		1		1	Adequate	

Date verified by the Committee Members:

Member of LIC

Member of LIC

Member of LIC

Chairman of LIC